



Eurofins | Focus

Client Number 9550

Unit C1 4 Pacific Rise, Mount Wellington, Auckland, 1061

Telephone 09 525-0568

www.focusanalytics.co.nz

Authorised Representative

Mr Vikram Pathania
Quality Manager

Programme

Chemical Testing Laboratory

Accreditation Number 1308

Initial Accreditation Date 17 September 2018

Conformance Standard

ISO/IEC 17025:2017

General requirements for the competence of testing and calibration laboratories

Laboratory Services Summary

2.64 Asbestos

Key Technical Personnel

Mr Vikram Pathania	2.64(a)(b)(c)
Ms Rebecca Rawlings	2.64(a)(b)(c)
Mr Colin Wang	2.64(a)(b)(c)

Operations Manager
Authorisation:

Issue 28

Date: 11/05/23

Page 1 of 2



Eurofins | Focus
 Chemical Testing Laboratory
SCOPE OF ACCREDITATION

Accreditation Number 1308

2.64 Asbestos

(a) Identification

Bulk identification of asbestos using polarised light microscopy

AS 4964:2004 Method for the Qualitative Identification of Asbestos in Bulk Samples

(b) Fibre Counts

(Analysis only)

Determination of air-borne asbestos fibre counts on membrane filters by light microscopy


National Occupational Health & Safety Commission: 3003(2005) Guidance Note on the Membrane Filter Method for Estimating Airborne Asbestos Fibres (2nd Ed).

(c) Field Sites

Establishment and operation of field sites for the purposes of asbestos identification and fibre counting, in accordance with procedures listed in 2.64(a)(b).

References

AS Australian Standards

Operations Manager Authorisation:		Issue 28	Date: 11/05/23	Page 2 of 2
--------------------------------------	---	----------	----------------	-------------